



Gait Trainer Specifications

Compact Profile	Length	34.5 in (0.88 m)
	Width	27.0 in (0.69 m)
	Base inner width (closed)	21.5 in (0.53 m)
	Base Inner width (open)	32.0 in (0.81 m)
	Height	39.0 in (0.99 m)
Patient Size	Maximum patient weight	350 lbs. (159 kg)
	Maximum patient height	6' 7" (2.0 m)
Unweighting	Unweighting range	20-150 lbs. (10.0-68.0 kg)
Dynamic Sit-to-Stand from:	Wheel chair Therapy platform Select Hospital beds	
Mobility	Therapist propelled from front or rear Patient propelled	
Electric Castor Operation	Pushbutton electric castor operation with Wheels swivels locked (travels in a straight line) Wheel swivels unlocked for maximum mobility Wheels braked	
Touch Screen Interface with Statistics	Distance Walked (ft, m) Average Speed (ft / min, m / sec) Maximum Speed (ft / min, m / sec) Average Unweighting (lbs., kg) Walking Time (minutes) Session Time (minutes)	
Device Weight	180 lbs. (82 kg)	
Accessories	Adjustable armrest platforms	
Battery	2.5 hour recharge 6 hour run time	



Exosuit Specifications

Size Range	Unisex sizing Waist sizes 23 to 52 in (47 – 103 cm) Fits short, medium, & long leg lengths
Wearing by Patient	As simple to put on and take off as a pair of pants Side zippers allow putting on over shoes Accommodates catheter
Patient Setup	Easy attachment of patient in less than 5 minutes Two quick connect latches attach pants to gait trainer Quick connect air fitting
Security	Belt and back straps provide additional support and safety
Construction	Breathable stretch fabrics maintain flexibility Front zipper Knee straps for hands on therapy
Air Pressure	Produces the unweighting effect Pressure Range 0.05 to 0.85 psi. (3 to 44 mm Hg) Typical pressure of 50% unweighting 0.5 psi (26 mm Hg) Equivalent to Class I to Class III compression stockings
Laundering	Machine wash, Air dry



Certifications

IEC 60601-1, ed.3.1, (Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.

FDA Listing Number 3013697488

Underwriters Laboratories (UL) to IEC 60601- 1, Edition 3 Harmonized Medical Device Standard.

Lite Run, Inc.
1000 Westgate Dr.
St. Paul, MN 55114
Website: literun.com
Email: sales@literun.com